

# Thermo Scientific Syncronis Silica HPLC Columns

Consistent, predictable separations, Column after column, time after time

- Highly pure, high surface area silica
- Excellent reproducibility for normal phase chromatography
- · Rigorously tested to ensure quality



#### **Specifications**


Particle size	1.7 µm, 3 µm, 5 µm	Carbon load	-
Pore size	100 Å	Endcapped	_
Surface area	320 m <sup>2</sup> /g	<b>USP</b> classification	L3
pH range	2-8		

## High efficiency columns for normal phase chromatography

Unbonded, highly pure base deactivated silica media that is the backbone of the Thermo Scientific<sup>™</sup> Syncronis<sup>™</sup> range of columns. With a high surface area, Syncronis Silica HPLC columns

are a powerful and efficient tool for the chromatography of non-polar and moderately polar organic compounds by normal phase chromatography. For high speed, high efficiency separations, Syncronis Silica HPLC columns are available with 1.7  $\mu$ m particle size.

#### **Application: Cetirizine (USP)**



Parameter	USP Specification	Measured (6 replicate injections)
Tailing factor	< 2.0	1.05
%RSD peak area	< 2%	0.17%

#### Column: Syncronis Silica 5 µm, 250 mm x 4.6 mm


1. Cetirizine







#### Normal phase retention and selectivity



Column: Mobile phase: Flow rate Temperature: Detection:

Injection volume:

- 1. Toluene 2. Nitrobenzene
- 5. p-nitroaniline

#### Syncronis Silica 5 µm, 150 mm x 4.0 mm

iso-octane:ethanol:water (85:14.7:0.3) 1.25 mL/min 40 °C 254 nm

2 uL

- 3. o-nitroaniline
- 4. m-nitroaniline

### **Ordering Information**

Description	Particle Size	Length (mm)	2.1 mm ID	3.0 mm ID	4.0 mm ID	4.6 mm ID
		30	97002-032130	_	_	_
Syncronis Silica	1.7 μm	50	97002-052130	97002-053030	-	97002-054630
		100	97002-102130	97002-103030	-	_
		50	97003-052130	-	-	_
	3 µm	100	_	-	-	97003-104630
		150	_	-	-	97003-154630
	5 µm	30	97005-032130	97005-033030	97005-034030	97005-034630
		50	97005-052130	97005-053030	97005-054030	97005-054630
		100	97005-102130	97005-103030	97005-104030	97005-104630
		150	97005-152130	97005-153030	97005-154030	97005-154630
		250	97005-252130	97005-253030	97005-254030	97005-254630
Drop-in guard cartridges (4/pk)	5 μm	10	97005-012101	97005-013001	97005-014001	_
UNIGUARD Guard Cartridge Holder			852-00	852-00	850-00	-

#### Consistent, predictable separations: Column after column, time after time

Syncronis HPLC columns are manufactured, packed and tested in ISO 9001:2008 accredited facilities. Each lot of silica is tested for the physical properties of the silica support and only released for production if it meets the stringent test specifications.

Each bonded lot of chromatographic packing material is rigorously tested for primary and secondary interactions with the bonded phase. New, enhanced, automated packing methods drive consistency even further and every column is individually tested to ensure that it meets the required quality.

These extensive testing and quality control procedures ensure the delivery of a consistent product, column after column.

#### thermoscientific.com/Syncronis

© 2012 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific Inc. and its subsidiaries Specifications, terms and pricing are subject to change. Not all products are available in all countries. Please consult your local sales representative for details.

USA and Canada +1 800 332 3331 France +33 (0)1 60 92 48 34 Germany +49 (0) 2423 9431 20 or 21 **United Kingdom** +44 (0)1928 534110

Japan +81 3 5826 1615 PS20604\_E 09/12S

**China** +86 21 68654588 +86 10 84193588 +86 20 83145199 800 810 5118 India +91 22 6742 9494 +91 27 1766 2352 Australia 1 300 735 292 (free call domestic) New Zealand 0800 933 966 (free call domestic) **All Other Enquiries** +44 (0) 1928 534 050

Technical Support North America +1 800 332 3331

Outside North America +44 (0) 1928 534 440



Part of Thermo Fisher Scientific



